



EMF ASSESSMENT REPORT

EN IEC 62311:2020
EN50665:2017

Report Reference No.....: T2240505696-EMF

Compiled by

(position+printed name+signature)...: File administrators Lena Wen

Supervised by

(position+printed name+signature)...: Technique principal Hugo Chen

Approved by

(position+printed name+signature)...: Manager Andy Zhang

Date of issue.....: 2024/5/23



Testing Laboratory Name.....: Shenzhen Tongzhou Testing Co.,Ltd

Address: 1th Floor, Building 1, Haomai High-tech Park, Huating Road 387,
Dalang Street, Longhua, Shenzhen, China

Applicant's name: **Jiangxi Small Walrus Intelligent Technology Co.,LTD**

Address: Factory No. 1, Hengsi Road, Qianping Industrial Zone, Le 'an
Industrial Park, Fuzhou City, Jiangxi Province, China

Test specification

Standard: **EN IEC 62311:2020**
EN50665:2017

TRF Originator.....: Shenzhen Tongzhou Testing Co.,Ltd

Master TRF.....: Dated 2024-01

Shenzhen Tongzhou Testing Co.,Ltd All rights reserved.

This publication may be reproduced in whole or in part for non-commercial purposes as long as the Shenzhen Tongzhou Testing Co.,Ltd is acknowledged as copyright owner and source of the material. Shenzhen Tongzhou Testing Co.,Ltd takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

Test item description: Smart WiFi Dome Camera

Trade Mark



Model/Type reference.....: XHX-PT09-DW

Listed Models: XHX-PT10-DW, XHX-PT11-DW, XHX-PT12-DW, XHX-PT13-DW,
SC489-DWB2, SC489-DWB3, SC489-DWB4, HHK-5362X, HHK-
5368X, HHK-5369X, HHK-5370X, HHK-5371X, HHK-5350X, HHK-
5320X, HHK-5362Y, HHK-5368XY, HHK-5369Y, HHK-5370Y,
HHK-5371Y, HHK-5350Y, HHK-5320Y, HHK-5362Z, HHK-5368Z,
HHK-5369Z, HHK-5370Z, HHK-5371Z, HHK-5350G, HHK-5320G,
HHK-5562Y, HHK-5568XY, HHK-5569Y, HHK-5570Y, HHK-5571Y,
HHK-5550Y, HHK-5520Y, HHK-5562Z, HHK-5568Z, HHK-5569Z,
HHK-5570Z, HHK-5571Z, HHK-5550Z, HHK-5520Z

Hardware Version.....: V1.0

Software Version: V3.2.1X

Ratings: Input:DC 12V

Result.....: **PASS**

**EMF ASSESSMENT REPORT**

Test Report No. :	TZ240505696-EMF	2024/5/23
		Date of issue

Equipment under Test : Smart WiFi Dome Camera

Model /Type : XHX-PT09-DW

Listed Models : XHX-PT10-DW, XHX-PT11-DW, XHX-PT12-DW, XHX-PT13-DW, SC489-DWB2, SC489-DWB3, SC489-DWB4, HHK-5362X, HHK-5368X, HHK-5369X, HHK-5370X, HHK-5371X, HHK-5350X, HHK-5320X, HHK-5362Y, HHK-5368XY, HHK-5369Y, HHK-5370Y, HHK-5371Y, HHK-5350Y, HHK-5320Y, HHK-5362Z, HHK-5368Z, HHK-5369Z, HHK-5370Z, HHK-5371Z, HHK-5350G, HHK-5320G, HHK-5562Y, HHK-5568XY, HHK-5569Y, HHK-5570Y, HHK-5571Y, HHK-5550Y, HHK-5520Y, HHK-5562Z, HHK-5568Z, HHK-5569Z, HHK-5570Z, HHK-5571Z, HHK-5550Z, HHK-5520Z

Applicant : Jiangxi Small Walrus Intelligent Technology Co.,LTD

Address : Factory No. 1, Hengsi Road, Qianping Industrial Zone, Le 'an Industrial Park, Fuzhou City, Jiangxi Province, China

Manufacturer : Jiangxi Small Walrus Intelligent Technology Co.,LTD

Address : Factory No. 1, Hengsi Road, Qianping Industrial Zone, Le 'an Industrial Park, Fuzhou City, Jiangxi Province, China

Test Result:	PASS
---------------------	-------------

The test report merely corresponds to the test sample.

It is not permitted to copy extracts of these test result without the written permission of the test laboratory.



Contents

<u>1.</u>	<u>TEST ENVIRONMENT</u>	<u>4</u>
1.1.	Address of the test laboratory	4
1.2.	Environmental conditions	4
1.3.	Statement of the measurement uncertainty	4
<u>2.</u>	<u>METHOD OF MEASUREMENT</u>	<u>5</u>
2.1.	Applicable Standard	5
2.2.	Introduction	5
2.3.	Limit	5
2.4.	EMF Assessment Method	6
<u>3.</u>	<u>TEST RESULT</u>	<u>6</u>
<u>4.</u>	<u>CONCLUSION</u>	<u>6</u>



1. TEST ENVIRONMENT

1.1. Address of the test laboratory

Shenzhen Tongzhou Testing Co.,Ltd
1th Floor, Building 1, Haomai High-tech Park, Huating Road 387, Dalang Street, Longhua, Shenzhen, China

The sites are constructed in conformance with the requirements of ANSI C63.7, ANSI C63.4 (2014) and CISPR Publication 22.

1.2. Environmental conditions

During the measurement the environmental conditions were within the listed ranges:

Temperature: 15-35 ° C

Humidity: 30-60 %

Atmospheric pressure: 950-1050mbar

1.3. Statement of the measurement uncertainty

The data and results referenced in this document are true and accurate. The reader is cautioned that there may be errors within the calibration limits of the equipment and facilities. The measurement uncertainty was calculated for all measurements listed in this test report acc. to TR-100028-01 "Electromagnetic compatibility and Radio spectrum Matters (ERM);Uncertainties in the measurement of mobile radio equipment characteristics; Part 1" and TR-100028-02 "Electromagnetic compatibility and Radio spectrum Matters (ERM);Uncertainties in the measurement of mobile radio equipment characteristics; Part 2 " and is documented in the Shenzhen Tongzhou Testing Co.,Ltd quality system acc. to DIN EN ISO/IEC 17025. Furthermore, component and process variability of devices similar to that tested may result in additional deviation. The manufacturer has the sole responsibility of continued compliance of the device.

Hereafter the best measurement capability for Shenzhen Tongzhou Testing Co.,Ltd is reported:

Test Items	Measurement Uncertainty	Notes
Transmitter power conducted	0.57 dB	(1)

(1) This uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=1.96.



2. Method of measurement

2.1. Applicable Standard

EN62311: Assessment of electronic and electrical equipment related to human exposure restrictions for electromagnetic fields (0 Hz – 300 GHz)

EN50665: Generic standard for assessment of electronic and electrical equipment related to human exposure restrictions for electromagnetic fields (0 Hz - 300 GHz)

Council Recommendation 1999/519/EC of 12 July 1999 on the limitation of exposure of the general public to electromagnetic fields (0Hz to 300GHz) (Official Journal L 197 of 30 July 1999).

2.2. Introduction

This International Standard provides simple conformity assessment methods for low-power electronic and electrical equipment to an exposure limit relevant to electromagnetic fields (EMF). If such equipment cannot be shown to comply with the applicable EMF exposure requirements using the methods included in this standard for EMF assessment, then other standards, including IEC 62311 or other (EMF) product standards, may be used for conformity assessment.

2.3. Limit

Basic restriction for electric, magnetic and electromagnetic fields (0Hz to 300GHz)

Frequency range	Magnetic flux density (mT)	Current density (mA/m ²)	Whole body average SAR(W/kg)	Localised SAR (head and trunk)(W/kg)	Localised SAR (limbs) (W/kg)	Power density, S (W/m ²)
0Hz	40	--	--	--	--	--
>0-1Hz	--	8	--	--	--	--
1-4Hz	--	8/f	--	--	--	--
4-1000Hz	--	2	--	--	--	--
1000Hz-100kHz	--	f/500	--	--	--	--
100kHz-10MHz	--	f/500	0.08	2	4	--
10MHz-10GHz	--	--	0.08	2	4	--
10-300GHz	--	--	--	--	--	10

Notes:

1. f is the frequency in Hz.
2. The basic restriction on the current density is intended to protect against acute exposure effects on central nervous system tissues in the head and trunk of the body and includes a safety factor. The basic restrictions for ELF fields are based on established adverse effects on the central nervous system. Such acute effects are essentially instantaneous and there is no scientific justification to modify the basic restrictions for exposure of short duration. However, since the basic restriction refers to adverse effects on the central nervous system, this basic restriction may permit higher current densities in body tissues other than the central nervous system under the same exposure conditions.
3. Because of electrical inhomogeneity of the body, current densities should be averaged over a cross section of 1cm² perpendicular to the current direction.
4. For frequencies up to 100 kHz, peak current density values can be obtained by multiplying the rms value by $\sqrt{2}$ (=1.414). For pulses of duration t_p the equivalent frequency to apply in the basic restrictions should be calculated as $=1/(2t_p)$
5. For frequencies up to 100 kHz and for pulsed magnetic fields, the maximum current density associated with the pulses can be calculated from the rise/fall times and the maximum rate of change of magnetic flux density. The induced current density can then be compared with the appropriate basic restriction.
6. All SAR values are to be averaged over any six-minute period.
7. Localised SAR averaging mass is any 10g of contiguous tissue; the maximum SAR so obtained should be the value used for the estimation of exposure. These 10g of tissue are intended to be a mass of contiguous tissue with nearly homogeneous electrical properties. In specifying a contiguous mass of tissue, it is recognised that this concept can be used in computational dosimeter but may present difficulties for direct physical measurements. A simple geometry such as cubic tissue mass can be used provided that the calculated dissymmetric quantities have conservation values relative to the exposure guidelines.
8. For pulses of duration t_p the equivalent frequency to apply in the basic restrictions should be calculated as $=1/(2t_p)$. Additionally, for pulsed exposures, in the frequency range 0.3 to 10GHz and for localised exposure



of the head, in order to limit and avoid auditory effects caused by thermoplastic expansion, an additional basic restriction is recommended. This is that SA should not exceed 2mJ kg⁻¹ averaged over 10g of tissue.

Reference levels for electric, magnetic and electromagnetic fields (0Hz to 300GHz, unperturbed rms values)

Frequency range	E-field strength (V/m)	H-field strength (A/m)	B-field (uT)	Equivalent plane wave power density $S_{eq}(W/m^2)$
0-1Hz	--	3.2×10^4	4×10^4	--
1-8Hz	10000	$3.2 \times 10^4 / f^2$	$4 \times 10^4 / f^2$	--
8-25Hz	10000	4000/f	5000/f	--
0.025-0.8KHz	250/f	4/f	5/f	--
0.8-3KHz	250/f	5	6.25	--
3-150KHz	87	5	6.25	--
0.15-1MHz	87	0.73/f	0.92/f	--
1-10MHz	$87/f^{1/2}$	0.73/f	0.92/f	--
10-400MHz	28	0.073	0.092	2
400-2000MHz	$1.375f^{1/2}$	$0.0037f^{1/2}$	$0.0046f^{1/2}$	f/200
2-300GHz	61	0.16	0.20	10

Notes: 1. As indicated in the frequency range column.

2. For frequencies between 100 kHz and 10GHz, S_{eq} , E^2 , H^2 and B^2 are to be averaged over any six-minute period.

3. For frequencies exceeding 10GHz, S_{eq} , E^2 , H^2 and B^2 are to be averaged over any $68/f^{1.05}$ -minute period (.in GHz).

4. No E-field value is provided for frequencies <1Hz, which are effectively static electric fields. For most people the annoying perception of surface electric charges will not occur at field strengths less than 20kV/m. Spark discharges causing stress or annoyance should be avoided.

2.4. EMF Assessment Method

Predication of EMF limit at a given distance

Equation from page 26 of EN 62311

$$E = \eta_0 H = \frac{\sqrt{30PG(\theta, \phi)}}{r}$$

Where:

E: E-field strength (V/m)

P: power input to antenna (Watt)

G: is the antenna gain relative to an isotropic antenna;

θ, ϕ : are elevation and azimuth angles to point of investigation;

r: is the distance from observation point to the antenna;

η_0 : is the characteristic impedance of free space.

As declared by the Applicant, the EUT transmits with the maximum source-baed Duty Cycle of 100%, and the EUT is a wireless device used in a medical application, at least 20cm from any body part of the user or nearby persons; from the maximum EUT RF output power, the minimum separation distance, $r=20$ cm, as well as the gain of the used antenna, the RF power density can be obtained.

3. Test Result

Type	Maximum Tune-up Conducted Output Power (dBm)	Antenna Gain(dBi)	Maximum EIRP (dBm)	Maximum Output Power(W)	Minimum Separation Distance(cm)	E-field strength cm(V/m)	E-field strength Limit(V/m)
WiFi2.4G	--	--	16.63	0.0460	20	5.9	61
WiFi5.2G	--	--	15.68	0.0370	20	5.3	61
WiFi5.8G	--	--	11.92	0.0156	20	3.4	61

4. Conclusion

The measurement results comply with the relevant limits for general public exposure specified as reference levels in the Council Recommendation 1999/519/EC.

.....End of Report.....